



The Royal Australian
and New Zealand
College of
Obstetricians and
Gynaecologists



Position statement on midurethral slings

This statement has been developed by the Royal Australian and New Zealand College of Obstetricians and Gynaecologists (RANZCOG) and the Urogynaecological Society of Australasia (UGSA).

Disclaimer This information is intended to provide general advice to practitioners. This information should not be relied on as a substitute for proper assessment with respect to the particular circumstances of each case and the needs of any patient. This document reflects emerging clinical and scientific advances as of the date issued and is subject to change. The document has been prepared having regard to general circumstances.

First endorsed by RANZCOG: March 2014

Current: March 2014

Review due: March 2017

This position statement by the Royal Australian and New Zealand College of Obstetricians and Gynaecologists (RANZCOG) and the Urogynaecological Society of Australasia (UGSA) supports the use of mid-urethral slings (MUS) in the surgical management of female stress urinary incontinence (SUI). This is the type of urinary leakage associated with physical exertion, such as coughing, laughing and sneezing.

Stress urinary incontinence is an extremely common¹, burdensome and costly condition for women in Australasia, with a negative impact on a women's quality of life. Non-surgical measures such as pelvic floor muscle training and behavioural modifications are useful treatment options in alleviating symptoms in some women, although many proceed with surgery, which is a more effective treatment.²

Mid-urethral slings are minimally invasive procedures developed in the early 1990s to treat female stress urinary incontinence. These slings are narrow, synthetic polypropylene tapes that are surgically placed beneath the middle part of the urethra (water pipe) to provide dynamic support to stop leakage from the bladder. They have been shown to be as effective as more invasive traditional surgery with major advantages of shorter operating and admission times, and a quicker return to normal activities, together with lower rates of complications.³ This has resulted in MUS becoming the operation of choice in Europe, the United Kingdom, Australasia⁴ and the USA⁵ for treatment of SUI.

The USA Food and Drug Administration (FDA) released a white paper⁶ and safety communications⁷ regarding safety and effectiveness of transvaginal placement of surgical mesh specifically for pelvic organ prolapse. A prolapse is where some of the pelvic organs bulge downwards giving rise to symptoms of an uncomfortable vaginal lump. Media attention⁸ on this totally distinct and separate issue of mesh use in women has the potential to cause unnecessary confusion and fear in women considering MUS for treatment of stress urinary incontinence. Both RANZCOG and UGSA wish to strongly emphasise that the US FDA publications clearly state that MUS were not the subject of their safety communication.

There is robust evidence⁹⁻¹¹ to support the use of MUS from over 2,000 publications making this treatment the most extensively reviewed and evaluated procedure for female stress urinary incontinence now in use. These scientific publications studied all types of patients, including those with co-morbidities such as prolapse, obesity and other types of bladder dysfunction. It is, however, acknowledged that any operation can cause complications and for MUS, these include bleeding, damage to the bladder and voiding difficulties¹². Nevertheless, the results of a recent large multi-centre trial¹³ have again confirmed the excellent outcomes and low risks of complications to be expected after treatment with MUS. Additionally, long term effectiveness has been demonstrated in studies following patients for up to 17 years.¹⁴⁻¹⁵ In Australia, it has been the operation of choice to treat for female SUI since 2004. RANZCOG and UGSA support the use of monofilament polypropylene mid-urethral sling for surgical treatment of female stress urinary incontinence.

References

1. <http://www.aihw.gov.au/publication-detail/?id=60129543605> [accessed 23Feb14]
2. Labrie J, Berghmanns BL, Fischer K et al, Surgery versus physiotherapy for stress urinary incontinence. *N Engl J Med*, 2013. 369(12): p. 1124-33. doi: 10.1056/NEJMoa1210627. PMID 24047061
3. Cody J, Wyness L, Wallace S et al. Systematic review of the clinical effectiveness and cost-effectiveness of tension-free vaginal tape for treatment of urinary stress incontinence. *Health Technol Assess* 2003; 7 (21): iii, 1–189.
4. Lee J, Dwyer PL. Age related trends in female Stress Urinary Incontinence Surgery in Australia – Medicare data 94 – 09. *Aust N Z J Obstet Gynaecol* 2010; 50: 543 - 549. doi:10.1111/j.1479-828X.2010.01217.x PMID:21133865
5. <http://www.augs.org/d/do/2535> [accessed 25feb14]
6. FDA, Urogynecologic Surgical Mesh: Update on the Safety and Effectiveness of Vaginal Placement for Pelvic Organ Prolapse. 2011: <http://www.fda.gov/downloads/medicaldevices/safety/alertsandnotices/UCM262760.pdf> .
7. FDA, FDA Safety Communication: UPDATE on Serious Complications Associated with Transvaginal Placement of Surgical Mesh for Pelvic Organ Prolapse <http://www.fda.gov/MedicalDevices/Safety/AlertsandNotices/ucm262435.htm> . 2011.
8. <http://www.bjuinternational.com/bjui-blog/midurethral-tape-surgery-for-incontinence-a-possible-victim-of-the-vaginal-mesh-crisis/> [accessed 23Feb14]
9. Ogah J, Cody JD, & Rogerson L. Minimally invasive synthetic suburethral sling operations for stress urinary incontinence in women. *Cochrane Database Syst. Rev.* CD006375 (2009).
10. Rehman H, Bezerra CC, Bruschini H, Cody JD. Traditional suburethral sling operations for urinary incontinence in women. *Cochrane Database Syst. Rev.* CD001754 (2011).
11. Novara, G., et al., Updated systematic review and meta-analysis of the comparative data on colposuspensions, pubovaginal slings, and midurethral tapes in the surgical treatment of female stress urinary incontinence. *Eur Urol*, 2010. 58(2): p. 218-38.
12. <http://www.mhra.gov.uk/Safetyinformation/Generalsafetyinformationandadvice/Product-specificinformationandadvice/Product-specificinformationandadvice%E2%80%933M%E2%80%933T/Syntheticvaginaltapesforstressincontinence/Summariesofthesafetyadverseeffectsofvaginaltapeslingsmeshesforstressurinaryincontinence/index.htm> [accessed 25Feb14]
13. Richter H E et al. Retropubic versus transobturator midurethral slings for stress incontinence. *N. Engl. J. Med.* 362, 2066–2076 (2010).
14. Nilsson CG, et al., Seventeen years' follow-up of the tension-free vaginal tape procedure for female stress urinary incontinence. *Int Urogynecol J*, 2013. 24(8): p. 1265-9.
15. Liapis A, Bakas P, Creatsas G. Long-term efficacy of tension-free vaginal tape in the management of stress urinary incontinence in women: efficacy at 5- and 7-year follow-up. *Int Urogynecol J Pelvic Floor Dysfunct* 2008; 19:150

Appendices

Appendix A Full Disclaimer

This information is intended to provide general advice to practitioners, and should not be relied on as a substitute for proper assessment with respect to the particular circumstances of each case and the needs of any patient.

This information has been prepared having regard to general circumstances. It is the responsibility of each practitioner to have regard to the particular circumstances of each case. Clinical management should be responsive to the needs of the individual patient and the particular circumstances of each case.

This information has been prepared having regard to the information available at the time of its preparation, and each practitioner should have regard to relevant information, research or material which may have been published or become available subsequently.

Whilst the College endeavours to ensure that information is accurate and current at the time of preparation, it takes no responsibility for matters arising from changed circumstances or information or material that may have become subsequently available.